

Serum Volatiles (Control) on MEGA-ALC 1 Column

Column: **MEGA-ALC 1 - 0.20mm, 1.10µm, 10m**
Catalog Code: C-ALC1-020-110-10

Conditions:

Injection: Split 250°C, 0.2µL, 100mL/min Split flow.

Sample: BIO-RAD (Ref. #: 384) Level 2 Liquichek
Serum Volatiles Control.

Detector: FID 250°C.

Oven Temp.: 35°C (0.5min), 10°C/min, 70°C.

Carrier Gas: Helium, 60 kPa.

Peaks:

1. Methanol
2. Ethanol
3. Acetone
4. Isopropanol
5. Ethylene Glycol

